



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

TSI Incorporated

**500 Cardigan Road
Shoreview, MN 55126**

Fulfils the requirements of

ISO/IEC 17025:2017

and national standard

ANSI/NCSL Z540-1-1994 (R2002)

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.

The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 20 February 2028

Certificate Number: AC-2850



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

TSI Incorporated

500 Cardigan Road
Shoreview, MN 55126

answers@tsi.com (+1) 800-680-1220

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **22 December 2025**

Certificate Number: **AC-2850** Certificate Expiry Date: **20 February 2028**

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Airborne particle counting efficiency ¹	(0.50 to 0.80) efficiency Particle size 10 nm > 0.90 efficiency Particle size 15 nm (0.38 to 0.62) efficiency Particle size 23 nm > 0.90 efficiency Particle size 41 nm (0.90 to 1.1) efficiency Particle size 55 nm	0.068 0.05 0.054 0.19 0.046	Electrometer, 3068B ISO 27891:2015
Airborne particle concentration counting efficiency ^{1,2}	(0.9 to 1.1) efficiency Particle Concentration Range 300 counts/cm ³ 600 counts/cm ³ 1 000 counts/cm ³ 2 000 counts/cm ³ 4 000 counts/cm ³ 6 000 counts/cm ³ 8 000 counts/cm ³ 10 000 counts/cm ³ 25 000 counts/cm ³ 50 000 counts/cm ³	0.11 0.11 0.13 0.03 0.04 0.03 0.03 0.03 0.04 0.04	Electrometer, 3068B ISO 27891:2015
Calibration factor for condensation particle counters (CPC/PNC)			

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Airborne particle counting efficiency ¹	(0.30 to 0.70) efficiency Particle size 0.1 μm Particle size 0.15 μm Particle size 0.2 μm Particle size 0.3 μm Particle size 0.5 μm (0.90 to 1.1) efficiency Particle size 1.0 μm	0.079 0.078 0.069 0.067 0.067 0.074	Calibration performed using monodispersed spherical particles method ThermoFisher Scientific Polystyrene Latex Microspheres ISO 21501-4:2018/Amd 1:2023, TSI 3068B
Threshold determination for spherical particles, size error	0.1 μm 0.15 μm 0.2 μm 0.25 μm 0.3 μm 0.5 μm 0.7 μm 1.0 μm 2.0 μm 3.0 μm 5.0 μm 10 μm 25 μm	0.004 μm 0.005 μm 0.007 4 μm 0.006 1 μm 0.006 1 μm 0.007 μm 0.013 μm 0.019 μm 0.02 μm 0.11 μm 0.061 μm 0.5 μm 1.2 μm	ISO 21501-4:2018/ AMD1:2023 ThermoFisher Scientific™ Polystyrene Latex Microspheres (“PSL”)

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pneumatic Differential Pressure (Anemometer)	(0 to 15) inH ₂ O	0.21 % of reading + 0.003 1 inH ₂ O	Comparison to MKS Pressure Transducer 220DD-00100A2B
Pneumatic Barometric Pressure (Anemometer)	(8 to 40) inHg	0.042 inHg	Comparison to Setra 276 Barometric Pressure Sensor

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Velocity	(35 to 8 000) fpm	2.6 % of reading	Comparison to MKS Instruments Pressure Transducer 220DD-00010A2B MKS Pressure Transducer-220DD-22769 Omega Thermistor ON-901-44030
Pneumatic Barometric Pressure (Flow Meter, Low & High)	14.2 psia 22.2 psia	0.051 psi 0.026 psi	Comparison to PPC4-ui A1.4 Ms/A200Kp Pressure Controller; Setra 276 Pressure Transducer
Pneumatic Differential Pressure (Flow Meter, Low)	(-153 to 153) cmH2O	0.21 % of reading	Comparison to PPC4-ui A1.4 Ms/A200Kp Pressure Controller
Pneumatic Differential Pressure (Flow Meter, Low)	(-11 to 152) psig	0.53 % of reading	Comparison to PPC4-ui A1.4 Ms/A200Kp Pressure Controller
Mass Flow (Gas Type: Air, O ₂ , N ₂)	(0.01 to 0.02) slpm (0.021 to 0.03) slpm (0.031 to 0.1) slpm	6.1 % of reading 4.1 % of reading 2.4 % of reading	Comparison to Flow Calibrator with FPP T-916-TD Bronkhorst Mercury Sealed Piston Prover
Mass Flow (Gas Type: Air, O ₂ , N ₂)	(0.11 to 0.2) slpm (0.21 to 0.4) slpm (0.41 to 0.8) slpm (0.81 to 1.6) slpm (1.61 to 3) slpm	2 % of reading 1.5 % of reading 1.2 % of reading 1.1 % of reading 1 % of reading	Comparison to Flow Calibrator with FPP T-950-TD Bronkhorst Mercury Sealed Piston Prover
Mass Flow (Gas Type: Air, O ₂ , N ₂)	(3 to 300) slpm	0.81 % of reading	Comparison to Flow Calibrator with Fluke (0.019, 0.039, 0.078) inch Sonic Nozzles
Mass Flow (Gas Type: CO ₂)	(0.01 to 0.02) slpm (0.021 to 0.03) slpm (0.031 to 0.1) slpm	5.9 % of reading 4 % of reading 2.4 % of reading	Comparison to Flow Calibrator with FPP T-916-TD Bronkhorst Mercury Sealed Piston Prover
	(0.11 to 0.2) slpm (0.21 to 0.4) slpm (0.41 to 0.8) slpm (0.81 to 1.6) slpm (1.61 to 3) slpm	2.3 % of reading 1.7 % of reading 1.3 % of reading 1.1 % of reading 1 % of reading	Comparison to Flow Calibrator with FPP T-950-TD Bronkhorst Mercury Sealed Piston Prover

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Mass Flow (Gas Type: CO ₂)	(3 to 50) slpm	0.76 % of reading	Comparison to Flow Calibrator with Fluke (0.019, 0.039) inch Sonic Nozzles
Mass Flow (Gas Type: N ₂ O)	(0.01 to 0.02) slpm (0.021 to 0.03) slpm (0.031 to 0.1) slpm	5.9 % of reading 4 % of reading 2.4 % of reading	Comparison to Flow Calibrator with FPP T-916-TD Bronkhorst Mercury Sealed Piston Prover
Mass Flow (Gas Type: N ₂ O)	(0.11 to 0.2) slpm (0.21 to 0.4) slpm (0.41 to 0.8) slpm (0.81 to 1.6) slpm (1.61 to 3) slpm	2.3 % of reading 1.7 % of reading 1.2 % of reading 1.1 % of reading 1 % of reading	Comparison to Flow Calibrator with FPP T-950-TD Bronkhorst Mercury Sealed Piston Prover
Mass Flow (Gas Type: N ₂ O)	(3 to 25) slpm	1.5 % of reading	Comparison to Flow Calibrator with Fluke (0.019, 0.039) inch Sonic Nozzles
Size Resolution	0.2 µm 0.3 µm 0.5 µm	0.007 4 µm 0.006 1 µm 0.007 µm	ISO 21501-4:2018/Amd 1:2023
Volume Flow (Gas Type: Air)	(2.83 to 100) slpm	1.5 % of reading	ISO 21501-4:2018/Amd 1:2023 using TSI Flow Meter
	1 slpm 28.3 slpm	3.4 % of reading 1.6 % of reading	ISO 21501-4:2018/Amd 1:2023 using Sensidyne Gilibrator
False Counts	(0 to 20) counts/m ³	3 counts	ISO 21501-4:2018/Amd 1:2023

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature	(0, 22, 60)°C	0.12 °C	Comparison to ThermoFisher Scientific Temperature Baths, Pt Resistance Thermistor


Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity	(9.8 to 95) %RH	0.61 %RH	Comparison to Thunder Scientific 2500 Humidity Chamber

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. Unitless linear measure.
2. The nominal values listed are approximate.



Jason Stine, Vice President

